

Certificate of Compliance

Certificate: 1216470

Master Contract: 204767

Project: 1216470

Date Issued: November 13, 2001


Issued to: Camille Bauer AG
Aarauerstrasse 7
CH-5610 Wohlen
SWITZERLAND

Attention: Mr. Achim Atland


The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by:


Enzo De Simone, B.A.Sc.

Authorized by:


Nick Alfano
Operations Manager

PRODUCTS

CLASS 3631 04 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 84 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - CERTIFIED TO U.S. STANDARDS

PART A: Power Measurement Transducers for single phase systems, Models SINEAX P530/Q531, fixed equipment, rated 24-60V dc/ac, 50/60Hz, 4.5VA or 85-230V dc/ac, 50/60Hz, 4.5VA, or powered from measuring inputs, Equipment Class II, Pollution Degree 2, Installation Category III.

PART B: Power Measurement Transducers for 3 phase systems, Models SINEAX P530/Q531, fixed equipment, rated 24-60V dc/ac, 50/60Hz, 4.5VA or 85-230V dc/ac, 50/60Hz, 4.5VA, or powered from measuring inputs, Equipment Class II, Pollution Degree 2, Installation Category III.

NOTE: The above models have the alternate model designation DW5-XXXX for Ohio Semitronics, Inc., where 'XXXX' may represent 2-3 digits and/or 1-4 alphanumeric symbols that indicate the input/output values.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 1010.1-92

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Includes Amendment 1)

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognised to perform certification to U.S. Standards.



CSA INTERNATIONAL

Certificate: 1216470

Master Contract: 204767

Project: 1216470

Date: November 13, 2001

-
- | | | |
|------------------------------|---|---|
| CAN/CSA-C22.2 No. 1010.1B-97 | - | Amendment 2 to CAN/CSA-C22.2 No. 1010.1-92, "Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements" |
| UL Std No. 3111.1 | - | Equipment for Measurement Use; Part 1: General Requirements |



Supplement to Certificate of Compliance

Certificate: 1216470

Master Contract: 204767

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
1216470	November 13, 2001	Original Certification.